





Negative Control

| Tube Name: control |         |         |                            |                        |                          |
|--------------------|---------|---------|----------------------------|------------------------|--------------------------|
| Population         | #Events | %Parent | control FITC-A<br>Geo Mean | control FITC-A<br>Mean | control FITC-A<br>Median |
| P1                 | 29,830  | 35.7    | ####                       | 353                    | 293                      |
| P2                 | 147     | 0.5     | 3,519                      | 11,296                 | 2,915                    |
| P3                 | 6       | 0.0     | 143,265                    | 163,485                | 161,738                  |



患者サンプル結果

| Tube Name: CD16 |         |         |                         |                     |                       |
|-----------------|---------|---------|-------------------------|---------------------|-----------------------|
| Population      | #Events | %Parent | CD16 FITC-A<br>Geo Mean | CD16 FITC-A<br>Mean | CD16 FITC-A<br>Median |
| P1              | 24,798  | 15.2    | ####                    | 96,790              | 102,041               |
| P2              | 24,615  | 99.3    | 85,445                  | 97,506              | 102,297               |
| P3              | 21,441  | 86.5    | 105,355                 | 108,056             | 106,809               |

P3 ゲートのGeo Mean 値で判定、大きく低下する場合はP2ゲート  
(カットオフ値<40000 IGDを強く疑う。<60000 IGDの可能性はある)